

CERTIFICATE OF COMPLIANCE

Certificate Number 20151118-E95701
Report Reference E95701-20151117
Issue Date 2015-NOVEMBER-18

Issued to: CONTA CLIP VERBINDUNGSTECHNIK GMBH
OTTO-HAHNSTRASSE 7
33161 HOVELHOF GERMANY

**This is to certify that
representative samples of**

COMPONENT - TERMINAL BLOCKS
Terminal Blocks, Cat. No(s). SRKD 2.5, SRKD 2.5/SV,
SRKD 4, SRKD 4/SV, SRKD 10, SRKD 10/SV.
Protective Conductor Terminal Blocks, Cat. No(s). SSLD
2.5, SSLD 4, SSLD 10.

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL1059, Terminal Blocks
CAN/CSA C22.2 No. 158-10, Terminal Blocks

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Recognized components are incomplete in certain constructional features or restricted in performance
capabilities and are intended for use as components of complete equipment submitted for investigation rather
than for direct separate installation in the field. The final acceptance of the component is dependent upon its
installation and use in complete equipment submitted to UL LLC.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please
contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

